

Timestrip neo Instructions

- 1 Activation** To activate, press and hold the button for >3.5 seconds until the green 'Power on' LED flashes 3 times.
- 2 Activation Indication** Once the indicator has been activated, the 'Power on' LED will flash every 8 seconds.
- 3 Start Delay** The device starts monitoring temperature after the 30-minute start delay.
- 4 No Breach** If no temperature breach has occurred, the green 'Power' LED will flash.
- 5 30 minutes Breach** When the temperature drops below 15°C for more than 60 minutes single event the red '15°C' LED will flash.
- 6 60 minutes Breach** When the temperature exceeds 25°C for more than 60 minutes single event, the red '25°C' LED will flash.

